

Delivering leading edge, innovative power solutions for over 30 years since 1984

10/24/2023

Product Change Notice # REV106191

Product Change Notice:

Update label specs to ISO 15223 medical device compliant label, and add " ICT/ " to ITE POWER SUPPLY, add Morocco, UKCA to approvals,
Remove note from deviation, change O/P ferrite to 17.5*28.5*12.5mm. update BLADE/CORD INCLUDED to FC-NA2-1830-1 with SPT-1, 2C

Products effected: PR91R2160T00-IMR6B With this PCN the spec has been revised to Rev. C

This change will have no effect on product's form/fit/function of Globtek product. For any questions about this product change notice, please contact your GlobTek Sales Engineer at (201) 784-1000, visit our website at www.globtek.com or Email: sales@globtek.com .